Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
PGDT	All Pilot +	Permobil & Quickie	PC045		В	These 2 are the same joystick module.
PGDT	Remote +	Pride Mobility Quantum	PC045		В	joystick module. Pride Mobility renamed theirs from Pilot + to Remote +
PGDT	R-Net	Permobil Amysystems	PC046		В	
PGDT	VSI (new)	Permobil	PC046		В	




Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
PGDT	VSI (older, 4 or 6 key)		PC046		В	
PGDT	VR2	Pride Mobility Quantum Quickie Amysystems	PC046		В	



Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
HMC International (Permobil NV)	CJA (Compact Joystick Advanced)	Permobil	PC047			
HMC International (Permobil NV)	Compact Joystick	Permobil Pride	PC047			
HMC International (Permobil NV)	Compact Joystick Lite	Permobil Pride	PC047			
HMC International (Permobil NV)	Mini Joystick					

Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
Dynamic Controls	LiNX LE Low DLX-REM050		PC047		A	
Dynamic Controls	LiNX LE Traditional DLX-REM060 DLX-REM100 & 110		PC047		A	
Dynamic Controls	LiNX 200 Series DLX-REM210 & 211	Invacare	PC047		A	
Dynamic Controls	LiNX 200 Series DLX-REM215 & 216		PC047		A	

Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
Dynamic Controls	LINX REM400 Series	Invacare	PC047		A	
Dynamic Controls	MK5 & 6i SPJ+ (Shark DK-REMD & DX2-REM420/421)	Invacare	PC046		A	Not to be confused with the old SPJ (not +), which was sold with the older Mark4/5 R2 electronics. Customer will need longer screws (10-32 x 3/4" or M5 x 19) and Ø3/8" x 5/16" thick spacers. (Ø9mm x 8mm)
Dynamic Controls	Linx DLX-REM500	Invacare	PC047		none	The PC047 adapter will need to be modified to work with the REM500. When mounting with the 30mm mounting holes there will be interference with the XLR socket.
Dynamic Controls	DX2-REM550/551	Invacare	PC046		A	



Bodypoint

Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
Dynamic Controls	MK5 MPJ	Invacare	PC046		A	PC046 holes line up, but will need special nuts to fit in the slots of the joystick module
Dynamic Controls	MK4 A, A+, RX	Invacare	PC042		A	
Dynamic Controls	MK4 PSF & PSR		PC042		A	
Dynamic Controls			PC046		A	
	MK4 TCHJOY & TCPDM4					

Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
Dynamic Controls		Invacare			A	The mounting hole pattern for this joystick module is unknown at this time.
	DX-REM34B					
Dynamic Controls		Invacare			A	The mounting hole pattern for this joystick module is unknown at this time.
	DX-REMG90					
Dynamic Controls	8	Invacare			A	The mounting hole pattern for this joystick module is unknown at this time.
ļ	A Series (same as Shark	?)				

Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
Curtis Instruments	Q-Logic NE & NE+	Pride Mobility Quantum	PC046		В	
Curtis Instruments	Q-Logic Remote & EX	Pride Mobility Quantum	PC046		В	
Curtis Instruments	enAble 40 series	Ottobock	PC046		В	
Curtis Instruments	enAble 40 series Q-Logic Stand-Alone	Pride Mobility Quantum			В	Bodypoint does not have an adapter to fit this joystick module at this time. The hole pattern on the bottom is 4 M4 holes on 48mm x 48mm spacing

Bodypoint

Manufacturer	Model	Used by Chair Manufacturer	Joystick Adapter	Joystick Module with Adapter	Stem Size A= 3/16" B= 1/4"	Comments
Sunrise	Mini Joystick					1 PC043 and 1 HW310-22 or 25 Frame Clamp can be used for mounting
	Compact Joystick	Invacare			A	Housing is Ø46mm (1-13/16")